

#### SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

#### CREAFORM INC. 4700 rue de la Pascaline St. Levis, QC, G6W 0L9 CANADA Martin Nadeau Phone: 418 833 4446

#### CALIBRATION

Valid To: June 30, 2025

Certificate Number: 4274.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

#### I. Dimensional

Parameter/Equipment	Range	$CMC^{2}(\pm)$	Comments
Optical Portable CMM – Volumetric Performance	Up to 2 m	10 µm	ATP902-01 (based on ASME B89.4.22-2004 except 5.2 using length bar standards
Optical Portable CMM –			ATP942-01 based on partial calibration per ISO 10360-12 using:
Probing Size Error	Dia: 50.8 mm	6 µm	Reference sphere
Length Measurement Error	Up to 2 m	12 µm	Length bar standards
Optical Portable 3D Scanner –			
Probing Size Error	Dia.: 38.1 mm	3 μm	ATP939-01 based on VDI/VDE 2634 Part 3
Sphere Spacing Error	Up to 1.5 m	12 μm	sections 4.1 and 4.2 using ball bar standards with reference spheres

(A2LA Cert. No. 4274.01) 06/16/2023

An

Parameter/Equipment	Range	$CMC^{2}(\pm)$	Comments
Portable 3D Scanner –			
Probing Size Error	Dia: 38.1 mm	4 μm	ATP927-01 based on VDI/VDE 2634 Part 3
Sphere Spacing Error	Up to 650 mm	12 μm	sections 4.1 and 4.2 using ball bar standards with reference spheres

<sup>1</sup> This laboratory offers commercial calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

An





# **Accredited Laboratory**

A2LA has accredited

## **CREAFORM INC.**

Levis, QC, CANADA

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This laboratory also meets the R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 16<sup>th</sup> day of June 2023.

Mr. Trace McInturff, Vice President, Accreditation Services For the Accreditation Council Certificate Number 4274.01 Valid to June 30, 2025

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.